

HYGISUN[®]

REF HS0602R



Medical Face Mask For Kid
Earloop. Blue

EN14683: 2019+AC: 2019 Type IIR

HYCISUN®

Medical Face Mask For Kid (Non-Sterile)

EARLOOP, BLUE, 145mmX95mm

Classified as Type IIR. Filtration efficiency (BFE) => >98%

ERSTE LAGE UNGEWOBENER STOFF

MITTLERE LAGE FILTERPAPIER (SCHMELZGEBLASENER STOFF)

DRITTE LAGE UNGEWOBENER STOFF

50 PCS



EN 14683: 2019+AC: 2019

REF HS0602A
10ER*5BAG



20 PCS



DRITTE LAGE UNGEWOBENER STOFF

MITTLERE LAGE FILTERPAPIER (SCHMELZGEBLASENER STOFF)

ERSTE LAGE UNGEWOBENER STOFF

Gebrauchsanleitung Instructions for use

Gebrauchsbestimmung / Intend use

Die Masken dienen als Schutzbarriere in einer nicht sterilen Umgebung der Gesundheitsvorsorge.
The masks are used to provide barrier protection in non-sterile healthcare environment.

Struktur und Material / Structure and Material

ungewobener PP-Stoff (Innen- und Außenschicht) mit Filterstoff (Mitte-Schicht), hitzegeformt.
PP Non-Woven fabric (Inner and outer layers) with filter fabric (Middle Layer) heat formed.

Vorsichtsmaßnahmen / Cautions

1. Überprüfen Sie die Vollständigkeit der Packung vor dem Gebrauch. Prüfen Sie das Label, Herstellungsdatum und die Haltbarkeitszeit um sicherzugehen, dass das Produkt innerhalb seiner Haltbarkeitszeit verwendet wird.
2. Nicht wiederverwenden, wenn die Verpackung beschädigt ist.
3. Nicht wiederverwenden, die Wiederverwendung kann zu Kreuzkontaminationen führen.
1. Check the package completeness before using. Check the label, manufacturing date and validity time, to make sure the product is in valid date.
2. Do not reuse if the package is damaged.
3. Do not reuse. Reusing may cause cross-contamination.

Lagerung / Storage

Das Produkt sollte in einer kühlen und trockenen Umgebung aufbewahrt werden, fern von Hitze und direkter Sonneneinstrahlung.
The product should be stored in a cool dry area, away from heat and direct sunlight.

Anleitung / How to wear



Ohrschlaufen über beide Ohren ziehen.
Ear strap on both ears.



Unteren Teil über das Kinn ziehen.
Pull down mask to cover jaw.



Nasenbrücke mit Maskenform anpassen.
Press down around nose.



HYCISUN®

CE EN 14683: 2019+AC: 2019

Medizinische Gesichtsmaske für Kinder (Nicht steril)

REF HS0602A
10ER*5BAG


OHRSCHLAUFE, BLAU, 145mmX95mm

Als Typ IIR klassifiziert; Filtrationseffizienz (BFE) = > 98%

(EN) Medical face mask
(FR) Masque facial médical
(NL) Medisch gezichtsmasker
(IT) Mascherina medica
(ES) Mascarina médica
(PT) Máscara facial médica

(SE) Medicinsk ansiktmask
(DK) Medicinsk ansigtmask
(NO) Medicinsk ansiktmaske
(HU) Orvosi arcmaszk
(CS) Lékařská obličejová maska
(HR) Medicinska maska za lice

50 Stk.



20 Stk.



(PT) Máscara facial médica
(IT) Mascherina medica
(ES) Mascarina médica
(FR) Masque facial médical

(HR) Medicinska maska za lice
(HU) Orvosi arcmaszk
(CS) Lékařská obličejová maska
(NO) Medicinsk ansiktmaske
(DK) Medicinsk ansigtmask
(SE) Medicinsk ansiktmask

Symbol	Bedeutung / Introductions	Symbol	Bedeutung / Introductions
	Nicht verwenden, wenn Verpackung defekt ist Don't use when packing damaged		"Nicht wiederverwenden" oder "Zur Einmal-Benutzung" "Don't reuse" or "Single use"
	Warnungen und Vorsichtsmaßnahmen Warnings and Precautions		Nicht essen Not edible
	Name und Adresse des Herstellers Manufacturer Name Address		Herstellungsdatum Manufacture Date
	Ohne Ethanol Alcohol free		Informationen des Vertreters der EU Informational European Union Representative
	Los-Code Batch Code		Haltbarkeit Shelf-life
	Bitte lesen und Anweisungen beachten Consult instructions for use		

 Sichuan Ai Doctor Medical Technology Co., Ltd.
333 Yongke Road, Yongsheng Town, Wenjiang District, Chengdu Sichuan, China

 Sunbeam International GmbH
Schumanstr. 12 52146, Würselen, Germany

 SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE Amsterdam, Netherlands


4 260676 530362

Hergestellt in China

 A20210305

 2021.03.05

 2023.03.04

Rev. 5/21 Rev. Date: 2021.02.12

Hergestellt in China


4 260676 530362

 2023.03.04

 2021.03.05

Medizinische Gesichtsmaske Für Kinder (Nicht steril)
Als Typ IIR klassifiziert; Filtrationseffizienz (BFE) => 98%

REF HS0602A
10ER*5BAG

HYCISUN®

Medical Face Mask For Kid (Non-Sterile)
Classified as Type IIR. Filtration efficiency (BFE) => 98%

50 Stk.

HYCISUN®
Medical Face Mask For Kid (Non-Sterile)
EAR LOOP, BLUE, 145mmX95mm
Classified as type IIR. Filtration efficiency (BFE) => 98%

SOFT LAYER UNTERKOPFWEICHE STREIFE
SOFT LAYER UNTERKOPFWEICHE STREIFE
SOFT LAYER UNTERKOPFWEICHE STREIFE

CE EN 14683: 2019+AC: 2019
REF HS0602A
10ER*5BAG

50 PCS



HYGISUN®



HYGISUN

EN14683: 2019+AC: 2019 TYPE IIR

Medizinische Gesichtsmaske für Kinder (Nicht steril)

OHRSCHLAUFE, BLAU, 145mm*95mm

Anzahl der Packstücke: 10

Als Typ IIR klassifiziert; Filtrationseffizienz (BFE) => 98%

- Produktname | HYGISUN Medizinische Gesichtsmaske für Kinder
- Modellspezifikation | Ohrschlaufen; blau; 145mmx95mm; nicht steril.
- Materialien | 70% Vliesstoffe, 30% schmelzgeblasene Vliesstoffe.
- Lagermethode | Masken sind in einem sauberen, belüfteten und trockenen aufzubewahren.
- Vorsichtsmaßnahmen | Es wird empfohlen, es alle 2-4 Stunden auszutauschen. Sobald es kontaminiert ist, sollte es so bald wie möglich ersetzt werden.

REF HS0802A
(10ER-PACKUNG)



4 260676 530379



Sichuan AI Doctor Medical Technology Co., Ltd.
333 Yongke Road, Yongsheng town, Wenjiang District,
Chengdu City, Sichuan Province, China

EC REP

SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE Amsterdam,
Netherlands.

LOT A20210305

2021.03.05

2023.03.04

HYGISUN®

Gebrauchsanleitung



[Gebrauchsbestimmung] Die Masken dienen als Schutzbarriere in einer nicht sterilen Umgebung der Gesundheitsvorsorge.
[Struktur und Material] ungewobener PP Stoff (Innen- und Außenlage) mit Filterstoff (Mittellage), hitzegeformt.

[Vorsichtsmaßnahmen]

- Überprüfen Sie die Vollständigkeit der Packung vor dem Gebrauch. Prüfen Sie das Label, Herstellungsdatum und die Haltbarkeitszeit um sicherzugehen, dass das Produkt innerhalb seiner Haltbarkeitszeit verwendet wird.
- Nicht benutzen, wenn die Verpackung beschädigt ist.
- Nicht wiederverwenden. Die Wiederverwendung kann zu Kreuzkontaminationen führen.

[Lagerung] Das Produkt sollte in einer kühlen und trockenen Umgebung aufbewahrt werden, fern von Hitze und direkter Sonnenstrahlung.



Anleitung



Ohrenschlaufen über beide Ohren hängen.



Nasensclipp an die Nasenform anpassen.



Unteren Teil über das Kinn ziehen.

HYGYNUN® EN14683:2019+AC:2019 TYPE IIR

Symbol	Bedeutung	Symbol	Bedeutung
	Nicht verwenden, wenn Verpackung defekt ist.		*Nicht wiederverwenden* oder *Zur Einmalbenutzung*
	Ohne Latex		Nicht steril
	Name und Adresse des Herstellers		Informationen des Vertreters der EU
	Los-Code		Herstellungsdatum
	Haltbarkeit		Vor direktem Sonnenlicht schützen
	Gebrauchsanleitung beachten		Trocken aufbewahren

HYGYNUN®

HYGYNUN® EN14683:2019+AC:2019 TYPE IIR

гласно стандарта EN14683:2019+AC:2019 TYPE IIR

гласно стандарта EN14683:2019+AC:2019 TYPE IIR

Symbol	Bedeutung	Symbol	Bedeutung
	Nicht verwenden, wenn Verpackung defekt ist.		*Nicht wiederverwenden* oder *Zur Einmalbenutzung*
	Ohne Latex		Nicht steril
	Name und Adresse des Herstellers		Informationen des Vertreters der EU
	Los-Code		Herstellungsdatum
	Haltbarkeit		Vor direktem Sonnenlicht schützen
	Gebrauchsanleitung beachten		Trocken aufbewahren



HYGISUN®



Medical Face Mask for Kid

Name: HYGISUN® Medical Face Mask for Kid

Model Specification: Earloop, Blue, 145mmX95mm, Non-Sterile.

Intended Use: The Mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids and particulate material. These Medical Face Masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single use, disposable device, provided non-sterile.

⚠ Cautions:

1. Check the package completeness before using. Check the label, manufacturing date and validity time, to make sure the product is in valid date.
2. Do not use if the package is damaged.
3. Do not reuse. Reusing may cause cross-contamination.

Instruction for use:

1. Open the packaging pouch and take out the mask.
2. Place the side with nose piece upward. Hang the ear loops on the ears.
3. Press the nose piece to fit the bridge of the nose, then press the nose piece and pull the lower end of the mask to the lower jaw.
4. Adjust the mask so that it covers the bridge of the nose to the lower jaw in order to get the best protection effect.

Storage:

The product should be stored in a cool dry place, away from heat and direct sunlight.

Shelf life: 2 Years

Manufacturer Information



Name: Sichuan Ai Doctor Medical Technology Co., Ltd.
Address: 333 Yongke Road, Yongsheng Town, Wengjiang District, Chengdu, Sichuan, China



Name: SUNGO Europe B.V.
Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands



Name: Sunbeam International GmbH
Address: Schumanstr. 12 52146, Würselen, Germany



HYGISUN®
 www.hygisun.de

Rev. S/ 21 Rev. Date: 2021.02.12



Symbols meaning:

Symbol	Introductions	Symbol	Introductions
	Batch-Code		Do not reuse/ single use/ Use only once
	Don't use when packing is damaged.		Manufacture Date
	Manufacturer Name and address		Name and Address of European Union Representative
	Latex Free		Shelf life
	Consult instructions for use		Importer information
	Warnings and Precautions		Non-sterile



Rev. S/ 21 Rev. Date: 2021.02.12

HYGISUN®
 www.hygisun.de

Rev. S/ 21 Rev. Date: 2021.02.12



Name: Sunbeam International GmbH
Address: Schumanstr. 12 52146, Würselen, Germany



Name: SUNGO Europe B.V.
Address: Olympisch Stadion 24, 1076DE Amsterdam, Netherlands

HYGISON®

Medical Face Mask For Kid (Non-Sterile)

EARLOOP, BLUE, 145mmX95mm
Classified as type IIR. Filtration efficiency (BFE) => >98%

ERSTE LAGE: UNGEWÖHNER STOFF
MITTLERE LAGE: FILTERPAPIER (SCHWELZBELASERER STOFF)
DRITTE LAGE: UNGEWÖHNER STOFF

50 pcs

CE EN 14683: 2019+AC: 2019

REF: HS0602A
10ER'SBAG



20 bC2

HYGISON®



Doc No: HS0602A-KID-CE-Q2

HYGISON®

Medical Face Mask For Kid

Name: HYGISON® Medical Face Mask for Kid
Model Description: Earloop, Blue, non-sterile, Non-Dyeable

This Mask is intended to be worn to protect both the patient and healthcare worker of microorganisms, body fluids and particulate matters. These masks are for use in infection control practices to reduce the potential of infection. This is a single use, disposable device, provided non-sterile.

Before use, please check the label, manufacturing date and the product is in valid date is damaged. Cause other contamination.

Do not touch the mask face when using. Keep away the bridge of the nose.

Dispose away.

HYGISON®

Disposal warning:

Disposal	Introduction	Disposal	Precautions
LOT	Search Code	<input checked="" type="checkbox"/>	Do not reuse after use! Use one once
<input checked="" type="checkbox"/>	Don't use when package is damaged	<input checked="" type="checkbox"/>	Manufacture Date
<input checked="" type="checkbox"/>	Manufacturer Name and address		
<input checked="" type="checkbox"/>	Label error		
<input checked="" type="checkbox"/>	Continue instructions for use		
	Warning and Precautions		



HYGISON®

Medizinische Gesichtsmaske für Kinder

CE EN14683: 2019+AC: 2019 TYPE IIR

Anzahl der Packstücke: 10

Als Typ IIR klassifiziert; Filtrationseffizienz (BFE) => >98%

1. Produktname & in welchem Zweckbereich verwendet wird
2. Produktbeschreibung (Eigenschaften, Bestandteile, Merkmale)
3. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
4. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
5. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
6. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
7. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
8. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
9. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)
10. Anzeigensystem (Eigenschaften, Bestandteile, Merkmale)

4 2604676 533079

HYGISON®

510mm X 346mm X 388mm

Medical Face Mask For Kid REF: HS0602A

Rev. S/21-1.0 Rev. Date 2021.02.12

All rights reserved by HYGISUN®

HYGISUN®

Medizinische Gesichtsmaske für Kinder

(NICHT STERIL)

REF HS0602R (10ER*5BAG)

OPP BAG EAN

BOX EAN



EN14683: 2019+AC: 2019 TYPE IIR

Hergestellt in China

Mass/Dimensions: 510mm X 346mm X 388mm
Menge: 40 Packungen (50Stk./Packung)
Quantity: 40 Box (50pcs/box)

IMPORTEUR

Sunbeam International GmbH
Schumanstr. 12 52146, Würselen, Germany



Sichuan Ai Doctor Medical Technology Co., Ltd.
333 Yongke Road, Yongsheng town, Wenjiang District, Chengdu, Sichuan, China

EC REP

SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE Amsterdam, Netherlands



HYGISUN®

MEDICAL FACE MASK FOR KID

(NON-STERILE)

REF HS0602R (10ER*5BAG)

OPP BAG EAN

BOX EAN



EN14683: 2019+AC: 2019 TYPE IIR

Made in China

Mass/Dimensions: 510mm X 346mm X 388mm
Menge: 40 Packungen (50Stk./Packung)
Quantity: 40 Box (50pcs/box)

IMPORTEUR

Sunbeam International GmbH
Schumanstr. 12 52146, Würselen, Germany



Sichuan Ai Doctor Medical Technology Co., Ltd.
333 Yongke Road, Yongsheng town, Wenjiang District, Chengdu, Sichuan, China

EC REP

SUNGO Europe B.V.
Olympisch Stadion 24, 1076DE Amsterdam, Netherlands





DECLARATION OF CONFORMITY
 ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

SUNGO Europe B.V.
 Olympisch Stadion 24, 1076DE
 Amsterdam, Netherlands
 SRN: NL-AR-000000247

Manufacturer

Name: SICHUAN AI DOCTOR MEDICAL
 TECHNOLOGY CO., LTD.
 Address: 333 Yongke Road, Yongsheng Town,
 Wenjiang District, Chengdu City, Sichuan Province,
 China

Conformity Assessment

Conformity Assessment Procedure
 Annex II+III of Regulation (EU) 2017/745

Applicable Standards

- EN ISO 14971: 2019
- EN ISO 15223-1: 2016
- EN 1041:2008+A1:2013
- ISO 10993-1: 2018
- EN ISO 10993-5: 2009
- EN ISO 10993-10: 2013
- EN 14683:2019+AC:2019 TYPE IIR

Product Information

Name: MEDICAL FACE MASK
 Model: HS0602A (145*95mm)
 GMDN: 35177
 Basic UDI-DI:
 Classification: Class I
 Rule: Rule 1, Annex VIII, Regulation (EU) 2017/745

Remark

The declaration of conformity is valid in connection
 with the release technical document XXXX.

All the supporting documentation is retained at the
 premises of the manufacturer.

The Declaration of Conformity is exclusively under
 the sole responsibility of the manufacturer.

On behalf of SUNGO Europe office, I confirmed we are
 EU REP of the company who issue this document.

Declaration

We herewith declare that the above-mentioned
 products meet the requirements of Medical Device
 Regulation (EU) 2017/745 and the applicable
 standards above.

Signature: *Ying Bo Tang* Date: March 15, 2021



Position: General Manager Place: Hefei /China



Authorized Signature (S)



> Retouradres Postbus 16114 2500 BC Den Haag

SUNGO Europe B.V.
T.a.v. de heer R. Luo
Olympisch Stadion 24
1076 DE Amsterdam

Datum: 19 maart 2021
Betreft: notificatie medisch hulpmiddel klasse I

Geachte heer Luo,

Hierbij bevestig ik de ontvangst op 15 maart 2021 van de notificatie van het medische hulpmiddel klasse I, dat bedrijf Sichuan Ai Doctor Medical Technology Co., Ltd., met Europees gemachtigde SUNGO Europe B.V., als fabrikant overeenkomstig Verordening (EU) 2017/745 (MDR) op de markt wenst te gaan brengen.

Het product is onder volgend kenmerk geregistreerd. Ik verzoek u om in alle verdere correspondentie over dit product het bijbehorende kenmerk te vermelden en het bij telefoongesprekken bij de hand te houden.

**Medical Face Mask
(geen merknaam) (NL-CA002-2021-56817)**

Ik wijs u erop dat medische hulpmiddelen die op de markt gebracht worden volgens de MDR over een systeem voor hulpmiddelindicatie (UDI) moeten beschikken¹ en dat fabrikanten, gemachtigden en importeurs in de Europese databank voor Europese hulpmiddelen (Eudamed) moeten worden geregistreerd². Bijlage VI van de MDR bevat de bij de registratie te verstrekken gegevens. Op dit moment is Eudamed nog niet in gebruik, zodat het wat betreft het bovenstaande voldoende is dat u uw product overeenkomstig de huidige wet- en regelgeving hebt genotificeerd.

Zodra Eudamed volledig in gebruik is, wordt de fabrikant of diens gemachtigde geacht binnen achttien maanden bovenstaand hulpmiddel te registreren in Eudamed.³

¹ O.g.v. art. 29 MDR.

² O.g.v. art. 31 MDR.

³ www.camd-europe.eu/wp-content/uploads/2018/05/FAQ_MDR_180117_V1.0-1.pdf. Zie vraag en antwoord nummer 20.

Farmatec

Bezoekadres:
Hoftoren
Rijnstraat 50
2515 XP Den Haag
T 070 340 6161

<http://hulpmiddelen.farmatec.nl>

Inlichtingen via:

medische_hulpmiddelen@
minvws.nl

Ons kenmerk:

CIBG-20211347

Bijlagen

-

Uw aanvraag


15 maart 2021

*Correspondentie uitsluitend
richten aan het retouradres met
vermelding van de datum en het
kenmerk van deze brief.*

Tevens wijs ik u er voor de goede orde nog op dat de registratie van uw mededeling betreffende de aflevering van het bovengenoemde product slechts een administratieve handeling betreft. Deze ontvangstbevestiging behelst dan ook geen besluit betreffende de kwalificatie van het desbetreffende product als medisch hulpmiddel in de zin van art. 1 WMH, noch betreffende de indeling in risicoklasse I.

De Minister voor Medische Zorg en Sport,
namens deze,

Afdelingshoofd
Farmatec

A handwritten signature in blue ink, consisting of a series of loops and a long horizontal stroke at the end.

Dr. M.J. van de Velde

NOTIS

SUNGO Europe B.V. | [Log out](#)

- > My cases
- > Notify medical devices
- > Change products
- > Apply for certificate of free sale

- > Find products

Notify Class I medical devices

General data

NOTIS number	20211347
Date of receipt	3/15/2021
Status	Finalised
Date finalised	3/24/2021

Client details

Name	SUNGO Europe B.V.
Contact person concerning this notification	Luo, dhr. R.

Manufacturer details

Authorised representative of manufacturer	Sichuan Ai Doctor Medical Technology Co., Ltd.
Address	No.333,Yongke Road, Yongsheng Town, Chengdu Cross-strait Science and Technology Industrial Development Park, Wenjiang District
Zipcode	611130
City	Chengdu, Sichuan
Country	CHINA

Products

Brand name	Group name	Article number(s)	Model(s)	Class	Status
	Medical Face Mask			MDR I	BEV

Documents

Date	File	Title
3/24/2021	20211347-0005.pdf	20211347
3/24/2021	20211347-0004.pdf	MDR Klasse I Bevestiging AR 1 product
3/15/2021	20211347-0001.pdf	Declaration of Conformity
3/15/2021	20211347-0002.docx	Product information

[Print](#) [Ok](#)



SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China

TÜV SÜD Products Testing (Shanghai) Co., Ltd.
B-3/4, No.1999 Du Hui Road, Minhang District
Shanghai 201108, P.R. China

CLIENT NAME Sichuan Ai Doctor Medical Technology Co., Ltd.

CLIENT ADDRESS 333 Yongke Road, Yongsheng Town, Wenjiang
District, Chengdu City, Sichuan Province, China.

TEST PERIOD 20-Mar-2021~28-Mar-2021

Prepared By


Bella Xu

(Bella Xu)
Report Drafter

Authorized By

Leo Liu

(Leo Liu)
Authorized Signatory



Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested.(3) The test report shall not be reproduced except in full without the written approval of the laboratory.(4) Without the agreement of the laboratory , the client is not authorized to use the test results for unapproved propaganda.

TEST REPORT

Sample Description : Medical Face Mask for Kid
Sample Quantity : 60 pieces
Lot Number/Batch Code : A20210305
Specification : /
Size : /
Brand Name : /
Style Number : HS0602A

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Method	Test Standard Type II R	Judgement
1	Bacterial Filtration Efficiency Test (BFE), %	EN 14683:2019+AC:2019(E) Annex B	≥ 98	Pass
2	Differential Pressure Test (Pa/cm ²)	EN 14683:2019+AC:2019(E) Annex C	< 60	Pass
3	Splash Resistance Pressure Test (kPa)	EN 14683:2019+AC:2019(E) ISO 22609:2004	≥ 16.0	Pass
4	Microbial Cleanliness Test (CFU/g)	EN 14683:2019+AC:2019(E) Annex D	≤ 30	Pass

Note: Pass = Meet customer requirements;
Fail = Fail customer requirements;
= No comment;
N.D. = Not detected.

Photo of Samples



Results

No.	Test Item	Test Result
1	Bacterial Filtration Efficiency (BFE) Test	Specimen 1#: 99.9% Specimen 2#: 99.8% Specimen 3#: 99.9% Specimen 4#: 99.9% Specimen 5#: 99.9%
2	Differential Pressure Test	31.8 Pa/cm ²
3	Splash Resistance Pressure Test	Specimen 1#~32#: None seen
4	Microbial Cleanliness Test	Specimen 1#: <1 CFU/g Specimen 2#: 7 CFU/g Specimen 3#: 3 CFU/g Specimen 4#: 12 CFU/g Specimen 5#: 7 CFU/g

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of masks.

2. Sample description was given by client

Sample description : Medical Face Mask for Kid
Specification : /
Lot Number : A20210305
Sample Receiving Date : 2021-03-20

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 *Staphylococcus aureus* ATCC 6538 (Particle Diameter 3.0±0.3µm).
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the culture in peptone water to achieve a concentration of approximately 5×10^5 CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specimen to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
 - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediately begin sampling the aerosol using the Anderson sampler.
 - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
 - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
 - 6.4.4 At the conclusion of the positive control run, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm²).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at (37±2)°C for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

$$BFE = (C - T) / C \times 100$$

T is the total plate count for the test specimen.

C is the mean of the total plate counts for the two positive controls.



8. Test results*

<i>P</i> Value Stage Number	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1	39	29	0	0	0	0	0	0
2	65	72	0	0	0	0	0	0
3	112	144	0	0	0	0	0	0
4	159	171	0	0	0	0	0	0
5	986	998	0	2	1	0	0	1
6	342	306	0	0	2	1	0	0
Total (<i>T</i>), CFU	1703	1720	<1	2	3	1	<1	1
Average (<i>C</i>), CFU	$1.7 \times 10^3 = (P_A + P_B) / 2$							
BFE, %				99.9	99.8	99.9	99.9	99.9
Requirements	≥ 98							
Remarks	<i>P</i> is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. <i>T</i> is the total of <i>P</i> value for the test specimen. <i>C</i> is the mean of the total of <i>P</i> value of the two positive controls.							



Differential pressure Test

1. Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Medical Face Mask for Kid
Specification : /
Lot Number : A20210305
Sample Receiving Date : 2021-03-20

3. Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5. Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5) °C and (85±5)% relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
- 6.2 The pretreated specimen is placed across the orifice (total area 4.9cm², test area diameter 25mm, airflow direction from the inside of the mask to the outside of the mask) and clamped into place so as to minimize air leaks.
- 6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
- 6.4 The differential pressure is read directly.
- 6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

Specimen	Test Results* (Pa/cm ²)	Average (Pa/cm ²)	Requirements	Judgement
1#	31.3	31.8	< 60	Pass
2#	31.9			
3#	31.2			
4#	32.7			
5#	32.0			

Splash Resistance Pressure Test

1. Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Medical Face Mask for Kid
Specification : /
Lot Number : A20210305
Sample Receiving Date : 2021-03-20

3. Test Method

EN 14683:2019+AC:2019(E).
ISO 22609:2004.

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5. Test specimen

- 5.1 As requested by client, take a total of 32 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at $(21\pm 5)^{\circ}\text{C}$ and $(85\pm 5)\%$ relative humidity.

6. Procedure

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).

- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)		
	Min.	Target	Max.
120	3.002	3.063	3.124

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within +2 % ~ -5 % of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula:
(p is the density of the test fluid.) $t = 0.5 + (2 \times p - g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} - g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.





Results:

Specimen	Test Results*	Requirements	Judgement
1#	None Seen	Pass Pressure at 16.0 kPa (120mmHg)	Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen		Pass
10#	None Seen		Pass
11#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass
14#	None Seen		Pass
15#	None Seen		Pass
16#	None Seen		Pass
17#	None Seen		Pass
18#	None Seen		Pass
19#	None Seen		Pass
20#	None Seen		Pass
21#	None Seen		Pass
22#	None Seen		Pass
23#	None Seen		Pass
24#	None Seen		Pass
25#	None Seen		Pass
26#	None Seen		Pass
27#	None Seen		Pass
28#	None Seen		Pass
29#	None Seen		Pass
30#	None Seen		Pass
31#	None Seen		Pass
32#	None Seen		Pass



Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Medical Face Mask for Kid
Specification : /
Lot Number : A20210305
Sample Receiving Date : 2021-03-20

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D.

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26)°C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by:

$$N_i = 3 n_i / M$$

$$\text{Microbial Cleanliness} = N_1 + N_2$$

$i = 1, 2.$

n = Colonies of the TSA plate or the SDA Plate.

M = Weight of the mask.


Results*:

Specimen	1#	2#	3#	4#	5#
Weight of the Mask (M, g)	2.56	2.57	2.59	2.60	2.61
Colonies of the TSA Plate (n ₁)	0	2	2	2	1
Colonies of the SDA Plate (n ₂)	0	3	0	7	4
Aerobic Microbial Number (N ₁ , CFU/g)	0	3	3	3	2
Fungi Number (N ₂ , CFU/g)	0	4	0	9	5
Microbial Cleanliness, (CFU/g)	<1	7	3	12	7
Requirements	≤ 30				

Note:

- *denotes this test was carried out by external laboratory assessed as competent.
- This report is for internal use only such as internal scientific research ,education, quality control, product R&D.

-END OF THE TEST REPORT-



SHAI) CO



中国认可
国际互认
检测
TESTING
CNAS L0412

检测报告

(Test Report)

No. GPLSR32Z566645L1

样品名称
(Sample Description)

一次性医用儿童口罩
Medical Face Mask for Kid

委托单位
(Applicant)

四川艾医生医疗科技有限公司
Sichuan Ai Doctor Medical Technology
Co.,Ltd.



声明 Statement

1. 本报告无检验检测专用章、报告骑缝章和批准人签章无效。
This report is invalid without special seal of inspection, cross-page seal and the approver's signatures.
2. 本报告页面所使用“PONY”、“谱尼”字样为本单位的注册商标，其受《中华人民共和国商标法》保护，任何未经本单位授权的擅自使用和仿冒、伪造、变造“PONY”、“谱尼”商标均为违法侵权行为，本单位将依法追究其法律责任。
The pattern and characters of "PONY" and "谱尼" used in this report are protected by the trademark law of the People's Republic of China. Any unauthorized usage, counterfeit, forgery and alteration of trademarks of "PONY" and "谱尼" are the violations of the law. The PONY has the right to pursue all legal liabilities of the subject of the delict.
3. 委托单位对报告数据如有异议，请于报告完成之日起十五日内（初级农产品报告请于报告收到之日起五日内）向本单位书面提出复测申请，同时附上报告原件并预付复测费。
If the applicant has any questions about the results, shall provide a written retest application with the original report, and prepay the retest fees to PONY within fifteen days since the approval date (as an exception, it shall be within five days since the date received for the primary agriculture products report).
4. 委托单位办理完毕以上手续后，本单位会尽快安排复测。如果复测结果与异议内容相符，本单位将退还委托单位的复测费。
After the applicant finishes the procedure mentioned above, PONY shall arrange the retest as soon as possible. If the retest result accords with the applicant dissent, PONY shall refund the retest fees.
5. 不可重复性或不能进行复测的实验，不进行复测，委托单位放弃异议权利。
Tests that can not be repeated and tested shall not be carried out again.
6. 委托单位对样品的代表性和资料的真实性负责，否则本单位不承担任何相关责任。
The applicant should undertake the responsibility for the provided samples' representativeness and document authenticity. Otherwise, PONY has not any relevant responsibilities.
7. 本报告仅对所测样品负责，报告数据仅反映对所测样品的评价，对于报告及所载内容的使用、使用所产生的直接或间接损失及一切法律后果，本单位不承担任何经济和法律后果。
This report is only responsible for the provided sample. The test results only represent the evaluation of the tested sample. PONY will not be responsible for any economical or legal liability generated from direct or indirect usage of the test report.
8. 本单位有权在完成报告后按规定方式处理所测样品。
PONY has the right to dispose the tested sample by rules, after approval of the test report.
9. 本单位保证工作的客观公正性，对委托单位的商业信息、技术文件等商业秘密履行保密义务。
PONY assures objectivity and impartiality of the test, and fulfills the obligation of confidentiality for applicant's commercial information, and technique document.
10. 本报告私自转让、盗用、冒用、涂改、未经本单位批准的复制（全文复制除外）或以其它任何形式的篡改均属无效，本单位将对上述行为追究其相应的法律责任。
The report is invalid in case of illegal transfer, embezzlement, imposture, modification or any altering, reproducing except in full, without approval of PONY. PONY shall investigate and affix the applicant's legal liability accordingly.

▲ 防伪说明 (Anti-counterfeiting Description):

- (1) 报告编号是唯一的;
The test report has exclusive report code.
- (2) 报告采用特制防伪纸张印制，纸张表面带有"PONY"防伪纹路，该防伪纹路不支持复印，即复制件不会带有"PONY"防伪纹路。
The test report is printed by anti-copying paper whose surface shows "PONY" security print with specific anticounterfeiting technique. Security print will disappear after copying. Duplicates are not expected to give "PONY" security print under any circumstances.



全国服务热线
400-819-5688

WWW.PONYTEST.COM

扫描二维码

关注谱尼测试微信

公众号 PONY4008195688



北京实验室：(010) 83055000
上海实验室：(021) 64851999
青岛实验室：(0532) 88706866
深圳实验室：(0755) 26050909
天津实验室：(022) 23607888
苏州实验室：(0512) 62997900

武汉实验室：(027) 85446975
武汉车附所：(027) 82318175
长春实验室：(0431) 80530198
大连实验室：(0411) 87336618
哈尔滨实验室：(0451) 58627755
郑州实验室：(0371) 69350670

新疆实验室：(0991) 6684186
石家庄实验室：(0311) 85376660
西安实验室：(029) 89608785
杭州实验室：(0571) 87219096
宁波实验室：(0574) 87977185
呼和浩特实验室：(0471) 3450025

太原实验室：(0351) 7555722
南宁实验室：(0771) 5518818
合肥实验室：(0551) 63843474
广州实验室：(020) 89224310
厦门实验室：(0592) 5568048
成都实验室：(028) 87702708

检测结果 (Test Results)

No. GPLSR32Z566645L1

第 1 页, 共 3 页 (page 1 of 3)

样品名称 (Sample Description)	一次性医用儿童口罩 Medical Face Mask for Kid	样品规格 (Sample Specification)	145*95MM
委托单位 (Applicant)	四川艾医生医疗科技有限公司 Sichuan Ai Doctor Medical Technology Co.,Ltd.	商标 (Trade Mark)	—
到样日期 (Received Date)	2021-03-05	生产日期或批号 (Manufacturing Date or Lot No.)	2021.02.26 20210226
检测日期 (Test Date)	2021-03-05~2021-03-17	样品等级 (Sample Grade)	—
样品状态 (Sample Status)	正常 Normal	检测类别 (Test Type)	委托检测 Commissioning Test
检测项目 (Test Items)	见下页 See next page	检测环境 (Test Environment)	符合要求 To meet the requirements
检测方法 (Test Methods)	见下页 See next page		
所用主要仪器 (Main Instruments)	口罩细菌过滤效率检测仪 等 Mask bacteria filtration efficiency detector etc.		
备注 (Note)	<p>1.型号: HS0602A Model: HS0602A</p> <p>2.生产单位/受检单位: 四川艾医生医疗科技有限公司 Manufacturer/ Tested company: Sichuan Ai Doctor Medical Technology Co.,Ltd.</p> <p>3.以上样品信息由委托单位提供 The information of sample was provided by the applicant</p> <p>4.该报告中检测方法由委托单位指定。 The testing methods mentioned in this report were designated by the applicant.</p> <p>5.限值标准: EN 14683:2019+AC:2019(E) (IIR 型) Limit Standard: EN 14683:2019+AC:2019(E) (Type IIR)</p>		
	编制人 (Edited by)	陈翔芳	
	审核人 (Checked by)	王明	
	批准人 (Approved by)	杨中崇	
	签发日期 (Issued Date)	2021 年 03 月 17 日	

检测结果

(Test Results)

No. GPLSR32Z566645L1

第 2 页, 共 3 页 (page 2 of 3)

序号 (S/N)	检测项目 (Test Item)	单位 (Unit)	限值 (Limit)	检测结果 (Test Result)			单项结论 (Evaluation)	检测方法 (Test Method)
1	细菌过滤效率 (BFE) Bacterial filtration efficiency(BFE)	%	≥98	99.95			符合 Pass	EN 14683:2019+ AC:2019(E) 附录 B Appendix B
				>99.99				
				99.82				
				99.95				
				99.91				
2	压力差 Differential pressure	Pa/cm ²	<60	A	B	C	符合 Pass	EN 14683:2019+ AC:2019(E) 附录 C Appendix C
				1-1	20.0	20.1		
				1-2	21.9			
				1-3	20.3			
				1-4	19.7			
				1-5	18.8			
				2-1	20.9	20.0		
				2-2	21.5			
				2-3	17.4			
				2-4	18.9			
				2-5	21.2			
				3-1	22.0	21.0		
				3-2	21.2			
				3-3	22.1			
				3-4	18.2			
				3-5	21.7			
				4-1	22.6	19.4		
				4-2	20.8			
				4-3	19.0			
				4-4	18.0			
				4-5	16.8			
				5-1	17.8	19.6		
				5-2	20.6			
				5-3	18.7			
				5-4	19.7			
5-5	21.1							

检测结果 (Test Results)

No. GPLSR32Z566645L1

第 3 页, 共 3 页 (page 3 of 3)

序号 (S/N)	检测项目 (Test Item)	单位 (Unit)	限值 (Limit)	检测结果 (Test Result)	单项结论 (Evaluation)	检测方法 (Test Method)
3	#抗溅压力 Splash resistance pressure	kPa	≥16.0	>16.0	符合 Pass	EN 14683:2019+ AC:2019
4	微生物洁净度 Microbial cleanliness	cfu/g	≤30	23	符合 Pass	EN 14683:2019+ AC:2019(E) 附录 D Appendix D
				19		
				12		
				20		
				18		

#表示为分包项目。


Sub-contracting items are marked with a “#” at its top-left corner.

承担分包单位: 谱尼测试集团上海有限公司 (资质认定证书编号 160920340809)

Subcontractor: Pony Testing International Group Shanghai Co., Ltd. (Accreditation certificate No. 160920340809)

A-试样编号-测试区域编号 Test Specimen number-Test area number; B-每个测试区域的压力差 Differential pressure for each test area; C-每个试样的平均压差 The averaged differential pressure for each test specimen.

照片 Photo:



——以下空白——
(End of Report)